



Test Report : GEM40I18

AC-DC High Reliability Interchangeable Medical Adaptor

■ DESIGN VERIFY TEST

Output Function Test

Input Function Test

Protection Function Test

■ SAFETY TEST

Safety Test

■ RELIABILITY TEST

Other test

DESIGN VERIFY TEST

OUTPUT FUNCTION TEST

| NO | TEST ITEM | SPECICATION | TEST CONDITION | RESULT | VERDICT |
|----|-------------------|-----------------|---|-----------------|---------|
| 1 | RIPPLE & NOISE | 120 mVp-p (Max) | I/P:230VAC O/P:FULL LOAD Ta:25°C | 62 mVp-p | P |
| 2 | VOLTAGE TOLERANCE | -2% ~ +2% (Max) | I/P:90VAC~264VAC O/P:FULL~MIN. LOAD Ta:25°C | -0.72% ~ +1.75% | P |
| 3 | LINE REGULATION | -1% ~ +1% (Max) | I/P:90VAC ~264VAC O/P:FULL LOAD Ta:25°C | -0% ~ +0.33% | P |
| 4 | LOAD REGULATION | -2% ~ +2% (Max) | I/P:230VAC O/P:FULL ~MIN LOAD Ta:25°C | -1.35% ~ +1.10% | P |
| 5 | SET UP TIME | 500 mS | I/P:230VAC O/P:FULL LOAD Ta:25°C | 248.5 mS | P |
| 6 | RISE TIME | 30 mS | I/P:230VAC O/P:FULL LOAD Ta:25°C | 55.9 mS | P |
| 7 | HOLD UP TIME | 16 mS (Min) | I/P:115VAC O/P:FULL LOAD Ta:25°C | 21.6 mS | P |

INPUT FUNCTION TEST

| NO | TEST ITEM | SPECICATION | TEST CONDITION | RESULT | VERDICT |
|----|---------------------------|--|--|------------------------------------|---------|
| 1 | VOLTAGE RANGE | 80VAC ~ 264VAC | I/P:TESTING O/P:FULL LOAD Ta:25°C | 61V ~ 264V | P |
| 2 | FREQUENCY RANGE | 47HZ - 63HZ (Typ) NO DAMAGE OSC | I/P: 100VAC ~ 240VAC O/P:FULL~MIN LOAD Ta:25°C | TEST: OK | P |
| 3 | EFFICIENCY | 88% | I/P:230VAC O/P:FULL LOAD Ta:25°C | 88.79% | P |
| 4 | AVERAGE EFFICIENCY | 87.59%(DoE Level VI) 88.60% (CoC Version 5) | I/P:115/230VAC O/P:25% 、 50% 、 75% 、 100% LOAD Ta:25°C | 88.89% (115VAC) 89.48% (230VAC) | P |
| 5 | AC CURRENT | 1.0 A (Max) | I/P: 100VAC O/P:FULL LOAD Ta:25°C | 0.848A | P |
| 6 | NO LOAD POWER CONSUMPTION | < 0.1W (Max) | I/P:230VAC O/P: NO LOAD Ta:25°C | 0.051W | P |

| | | | | | |
|---|-----------------|---------------------|--|--------------------------|---|
| 7 | INRUSH CURRENT | < 65A COLD START | I/P:230VAC O/P:FULL LOAD Ta:25°C | 65.5A | P |
| 8 | LEAKAGE CURRENT | < 100μA | I/P:264VAC O/P:Min LOAD Ta:25°C | L-FG: 20μA N-FG: 20μA | P |

PROTECTION FUNCTION TEST

| NO | TEST ITEM | SPECICATION | TEST CONDITION | RESULT | VERDICT |
|----|-------------------------|----------------------------------|--|--|---------|
| 1 | OVER LOAD PROTECTION | 105%~160% | I/P:230VAC O/P:TESTING Ta:25°C | 150% HICCUP MODE RESET : AUTO RECOVER | P |
| 2 | OVER VOLTAGE PROTECTION | 110%~140% | I/P:230VAC O/P:MIN LOAD Ta:25°C | 136% (1N4736A) Clamp by ZENER diode | P |
| 3 | SHORT PROTECTION | SHORT OUTPUT 1 HOUR NO DAMAGE | I/P:264VAC O/P:FULL LOAD Ta:25°C | NO DAMAGE HICCUP MODE RESET AUTO RECOVER | P |

■ SAFETY TEST

SAFETY TEST

| NO | TEST ITEM | SPECICATION | TEST CONDITION | RESULT | VERDICT |
|----|----------------------|----------------------|---------------------------------|------------------------------|---------|
| 1 | WITHSTAND VOLTAGE | I/P-O/P:5656 VDC/min | I/P-O/P:5656 VDC/min Ta:25°C | I/P-O/P: 0.03uA NO DAMAGE | P |
| 2 | ISOLATION RESISTANCE | I/P-O/P:500VDC>100MΩ | I/P-O/P:500 VDC Ta:25°C | I/P-O/P>100MΩ NO DAMAGE | P |

■ RELIABILITY TEST

OTHER

| NO | TEST ITEM | SPECICATION | TEST CONDITION | RESULT | VERDICT |
|----|-----------|--|----------------|--------|---------|
| 1 | MTBF | MIL-KDBK-217F NOTICES 2 PARTS COUNT TOTAL FAILURE RATE : 2.122442 M.T.B.F : 471155 HRS | | | P |

| TEST RESULT | TESTER | APPROVAL |
|-------------|-------------|---------------|
| PASS | PETER CHENG | VINCENT TSENG |