

**Business Stream Products
Certification Department**

TÜV Rheinland LGA Products GmbH · 90431 Nürnberg

Eljintek, Inc.
6F-1, No. 286-3, Shin Ya Road
CHIEN CHEN DIST., KAOHSIUNG 806
TAIWAN

Contact

Tel. +49 911 655-5225
Mail service@de.tuv.com

Date October 15, 2019

Application for : **CB-Zertifikat**
Certificate No. : DE 02026059
Device : External Switching PSU
Switching Power Supply
Type : see Certificate
Test requirement : IEC 60601-1:2005+A1
IEC 60601-1-6:2010+A1
IEC 60601-1-11:2015
for national deviations see test report

Dear Madame or Sir,

The submitted revision for the above mentioned product has been tested and in this configuration found to be in accordance with the above mentioned requirements.

Please present the attached certificate: no. DE 02026059 together with the basic certificate whenever you are asking for national approvals based on this certificate. Enclosed please find the certificate: no. DE 02026059

Kind regards

Certification body



Dipl.-Ing. (FH) A. Klinker

Test sample: no, documentation available

TÜV Rheinland
LGA Products GmbH

Tillystraße 2
90431 Nürnberg

Tel. +49 911 655-5225
Fax +49 911 655-5226
Mail service@de.tuv.com
Web www.tuv.com/safety

Board of Management

Dipl.-Ing.
Jörg Mähler, Spokesman

Dipl.-Kfm.
Dr. Jörg Schlösser

Chairman of the
Supervisory Board

Dipl.-Ing.
Ralf Scheller

Nuremberg HRB 26013
VAT No.: DE 811835490

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switching Power Supply
Name and address of the applicant	Eljintek, Inc. 6F-1, No. 286-3, Shin Ya Road Chien Chen Dist., Kaohsiung 806 Taiwan
Name and address of the manufacturer	Eljintek, Inc. 6F-1, No. 286-3, Shin Ya Road Chien Chen Dist., Kaohsiung 806 Taiwan
Name and address of the factory	(See appendix for factories information)
Ratings and principal characteristics	AC 100-240V; 50/60Hz; 0.18-0.09A; Class II; for output ratings refer to test report no. 50219325 004 page 10
Trademark (if any)	MW
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	GSM06xbzzzzz, GEM06lbzzzzz (x = E, U, SA, SAA, BS, KR or CH to denote for the type of plug; b = 05, 06, 07, 09, 12, 15, 18 or 24 to denote for the output ratings; z = 0 to 9, A to Z or blank for marketing purpose only)
Additional information (if necessary may also be reported on page 2)	-add new models with new trademark; for details see test report; -see also test report nos. 50219325 001-003
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1 IEC 60601-1-6:2010+A1 IEC 60601-1-11:2015 for national deviations see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	50219325 004

This CB Test Certificate is issued by the National Certification Body



TÜVRheinland®

TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com



Date: 15.10.2019

Signature:

Dipl.-Ing. (FH) A. Klumke

1. MEAN WELL Enterprises Co.,Ltd.
No. 28, Wuquan 3rd Rd.
Wugu District
New Taipei City 24891, Taiwan
2. Good Opportunity Electronic Co., Ltd.
Wan Hu Xi Rd, Shi Wan Cheng
Bo Luo Xian, Hui Zhou City
Guangdong, China
3. Youngyear Electronics Co., Ltd.
Jin Zhu Industries Area
Qing Xi Town, Dongguan City
Guandong
China
4. Eljintek, Inc.
6F-1, No. 286-3, Shin Ya Road
Chien Chen Dist., Kaohsiung 806
Taiwan
5. Mitra Power Solutions Co., Ltd.
First Industrial district
Cunwei Village, HengLi Town
Dong Guan, Guangdong, China

Additional information (if necessary)
Information complémentaire (si nécessaire)



Date: 15.10.2019

Signature:

Dipl.-Ing. (FH) A. Klinker