

Ref. Certif. No.

NL-105692

(IECEE) CB SCHEME **CB TEST CERTIFICATE** Product AC/DC Switching Adaptor Name and address of the applicant MEAN WELL Enterprises Co., Ltd. Name and address of the manufacturer MEAN WELL Enterprises Co., Ltd. Name and address of the factory SuZhou MEAN WELL Technology Co., Ltd.

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT

When more than one factory

Ratings and principal characteristics

Trademark / Brand (if any)

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

Additional information

A sample of the product was tested and found to be in conformity with

No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

No.269, Changping Road, Huangdai Town, Xiangcheng District, 215152 Suzhou Jiangsu, China

☑ Additional information on page 2

Product data Ratings: Input : 100-240Vac, 50/60Hz, Output: see test reports for details Product data - type NGE45xyzzzz Ratings: 1,0-0,6A Product data – type NGE65xyzzzz Ratings: 1,5-0,8A



NGE45xyzzzz and NGE65xyzzzz

Additional information on page 2

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, IEC 60601-1-11:2015 and IEC 60601-1-11:2015/AMD1:2020

National differences: AU, EU Group Differences, EU Special National Conditions, GB, JP, SA

Comments: This CB test certificate replaces CB test certificate NL-103811.

As shown in the Test Report Ref. No. which forms part of this Certificate

4915474.50, 4915474.51

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V. Meander 1051 6825 MJ Arnhem Netherlands



Signature: K. Lin

Page 1 of 2

Date: 2 April 2024





NL-105692

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Additional factory

MEAN WELL (GUANGZHOU) Electronics Co., Ltd. No.11 Jingu South Road, Huadu District, 510890 Guangzhou Guangdong, China

MEAN WELL Enterprises Co., Ltd. No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

YONGDEN TECHNOLOGY CORPORATION 345 MacArthur HighWay Tabang, 3015 Guiguinto, Bulacan, Philippines

MEAN WELL INDIA ELECTRONICS PRIVATE LIMITED 9C, Peenya industrial area, Chokkasandra, 2ND PHASE, PEENYA, 560058 Bengaluru (Bangalore) Urban, Karnataka India

Additional information

Model reference:

x=U,E,I,UK,AU,CN, KR, IN; y=05, 09, 12, 15, 18, 24, 48, zzzz=maybe Blank, -,0~9,A~Z or a~z for market purpose The risk management requirements were not addressed. The Usability evaluation has not been addressed.

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V. Meander 1051 6825 MJ Arnhem Netherlands



Signature: K. Lin

Date: 2 April 2024