



Declaration of Conformity

For the following equipment:

Product Name: Medical Type Switching Power Supply

Model Designation: RPS-200-12, RPS-200-15, RPS-200-24, RPS-200-27, RPS-200-48, RPS-200-X(X=12,15,24,27,48)

is herewith confirmed to comply with the requirements set out in the Council Directive, the following standards were applied:

RoHS Directive (2011/65/EU), (EU)2015/863

MDR Directive (EU) 2017/745

TUV certificate No: TA 50347614 for RPS-200-x-C TA 50348281 for RPS-200-x

EN 60601-1:2006+A1+A12+A2

MDR Directive (EU) 2017/745

EN 60601-1-2:2015

Conducted emission	EN 55011:2016+A11:2020	Class B
Radiated emission	EN 55011:2016+A11:2020	Class A(for Class II), Class B(for Class I)
Harmonic current	EN IEC 61000-3-2:2019+A1:2021	
Voltage flicker	EN 61000-3-3:2013+A1:2019+A2:2021	

EMS (Electro-Magnetic Susceptibility)

EMS (Electro-Magnetic Susceptibility)				
ESD air	EN 61000-4-2:2009	Level 4	15KV	
ESD contact	EN 61000-4-2:2009	Level 4	8KV	
RF field susceptibility	EN 61000-4-3:2020	Level 3	10V/m(80MHz-2.7GHz)	
RF field susceptibility	EN 61000-4-3:2020	Table 9	9~28V/m (385MHz~5.78GHz)	
EFT bursts	EN 61000-4-4:2012	Level 3	2KV/100KHz	
Surge susceptibility	EN 61000-4-5:2014+A1: 2017	Level 3	2KV/Line-Line	
Surge susceptibility	EN 61000-4-5:2014+A1: 2017	Level 4	4KV/Line-Earth	
Conducted susceptibility	EN 61000-4-6:2014	Level 3	10V	
Magnetic field immunity	EN 61000-4-8:2010	Level 4	30A/m	
·	EN IEC 61000-4-11:2020 09	%residualvoltage for 0.5	cycles,0% residual voltage for 1 cycles,	
Voltage dip, interruption	70% residual voltage for 25 cycles, 0% residual voltage for 250 cycles			

Note:

A component power supply with load will be installed into final equipment which consists of an electronically shielded metal enclosure. Since EMC performance will be affected by the complete installation, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again.

The EMC tests mentioned above are performed using a well defined metal plate to simulate said metal enclosure. For guidance on how to perform these EMC tests, please refer to "EMI testing of component power supplies".(as available on http://www.meanwell.com)" and TDF (Technical Documentation File).

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number SC3xxxxxxx

Person responsible for marking this declaration:

Mean Well Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D:

(Name / Position)

Taiwan

(Place)

(Signature)
Nov. 30th, 2023

Alex Tsai/Director, Product Strategy Center:

(Name / Position)

(Signature)

Version: 8